

MAYO HOSPITAL LAHORE

NO. / /MH

Dated Lahore the /2020

To.

Chief Operating officer,

Mayo Hospital, Lahore.


SUBJECT: TECHNICAL EVALUATION REPORT OF TECHNICAL BIDS FOR PROCUREMENT OF EYE DISPOSABLE ITEMS OA63 (YEAR2020-2021)


Respected Sir,


With reference to your office letter No.PC/58702-06/MH dated: 19-8-2020 ,regarding the above noted subject.


The technical scrutiny report and their qualifying status are as under pertaining to the following firms.


Sr.No	Name Of Firms	Status of Responsive Items	Status of Non-Responsive Items	REMARKS (Non-Responsive)
1.	Ali Gohar & Co		22.-----	No infusion. There is no compatibility.
2.	Optisurge	33.-----	04.---- 05.---	Part C not qualified .Knife is blunt Part C not qualified .Knife is blunt
3.	Optimed	02,03,04,05,09,10,18, 22,23,32-----	11 14, 15,	Part C not qualified .SNP Part C not qualified. Haptic are very soft and bend on holding Part A not qualified. FSC of Interjojo Inc. not provided
4.	Sohrab Jee Patel & Co	01,09,14,33.---	-----	
5.	Jasani Scientific	08,24,25,27,31.	07.---- 26.----	Tubing needed for cutter function is not available. Non autoclavable tubing set.
6.	Kamran Bashir & Brothers	04,05,06,12	-----	
7.	Medi serve international	01,03,04,05,06,10,12, 14,15,16,18,28	19 20	Part C not qualified. Gas level reduces very first day. Part C not qualified. Quality unsatisfactory
8.	Bussiness Medical International	12	14 16	Part A not qualified. Product quality certificates not provided. Part A&C not qualified. Product quality certificates and Samples not provided.
9.	Mubarak Vision	02,03,08(b),09,10,11, 17,18,19,20,22,23--- <b>NOTE:</b> For T.E.No. 07,08. Two samples were provided and company undertake to provide supply against item no 7b and 8b.(undertaking attached)	04,--- 05,--- 06,--- 07,--- 12,--- 15,--- 33---	Part C not qualified .Knife is blunt Part C not qualified .Knife is blunt Part C not qualified .Knife is blunt Part C not qualified .Cutting is not smooth, a lot of traction. Part C not qualified .Knife is blunt Part A not qualified .FSC of Interjojo Inc. not provided. Part-C not qualified. Albomed contains HPMC which has long chains of hydro-carbons. It causes raise in FOP and we cannot leave it in anterior chamber after surgery.

  
Mr. Muhammad Zayed Izhar  
Hospital pharmacist  
(Member)

  
Miss Amber Iqbal Khan  
Deputy Drug Controller  
(Member)

  
Dr. Khalid Mehmood Awan  
AMS (Admn-II)  
(Member)

  
Dr. Ali Ayaz  
Associate Professor Ophthalmology  
(Member)

  
Prof. Dr. Nasir Chaudhary  
Professor of Ophthalmology  
(Chairman)